

Contract Number: 800029

**Certificate: SGSNA/17/HEL/00068**  
Test Report Number: 287045-2

**Certified Product:** Pressure and optical products  
**Trademark (if any):** -  
**Model(s):** VOM, VOD, VOF, VOA, VCT, VCF, VCA, VCB, VCK, VCL, VG, VT, VDT, VDTL, VDU, VL, VVF  
**Technical Data:** 12 – 35 VDC, 50 mA max (VL, VVF, VG, VT, VDTL, VDT, VDU); 24 VDC, 200 mA max (VOM, VOD, VCA, VCB, VCF, VCT, VOF, VOA, VCK, VCL)



**Name and Address of Applicant:** Satron Instruments Inc.  
Lumpeenkatu 1, 33900, Tampere, Finland

**Name and Address of Manufacturer:** Satron Instruments Inc.  
Lumpeenkatu 1, 33900, Tampere, Finland

**Name and Address of Factory:** Satron Instruments Inc.  
Lumpeenkatu 1, 33900, Tampere, Finland

This certificate supercedes previous certificates issued with the same certificate number. Certification is valid when products are indicated on the SGS directory of certified products at [www.sgs.com](http://www.sgs.com). The product is certified according to ISO/IEC Guide 17067, *Conformity assessment - Fundamentals of product certification, System 3*, and in accordance with


**UL 61010-1, 3rd Ed. Rev. May 11, 2012**  
**CAN/CSA C22.2 No. 61010-1-12, 3rd Ed.**

Authorized by

Effective date: 20 September 2017

  
Paul Krauss  
Certifier

Consumer and Retail Services, a division of SGS North America Inc.  
620 Old Peachtree Road, Ste. 100, Suwanee, GA 30024, USA  
t +1 770 570 1800 f +1 770 277 1240 [www.sgs.com](http://www.sgs.com)  
Page 1 of 1



This certificate is issued by the company under its General Conditions for Certification Services accessible at [www.sgs.com/terms\\_and\\_conditions.htm](http://www.sgs.com/terms_and_conditions.htm). Attention is drawn to the limitations of liability defined therein and in the Test Report here above mentioned which findings are reflected in this Certificate. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.